


America

CERTIFICATE

No. U8 124422 0001 Rev. 00

Holder of Certificate: **HUE LIGHT CO., LTD.**
 E-1311, 60, Haan-ro
 Gwangmyeong-si, Gyeonggi-do 14322
 REPUBLIC OF KOREA

Certification Mark:




Product: **Control Equipment
 (Molecular Hydrogen Inhalation Device)**

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing, Certification, Validation and Verification Regulations (TCVVR)". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: CPSA07110824

Date, 2024-09-23

(Harry Kwon)



America

CERTIFICATE

No. U8 124422 0001 Rev. 00

Model(s): H-2000

Brand Name: Hue Light

Brand:



Tested according to:

UL 61010-1:2012/R:2023-06
CSA C22.2 No. 61010-1:2012/U3:2023-06

Parameters:

Rated input voltage: 100-240 VAC
Rated frequency: 50/60 Hz
Rated input current: 5 A
Protection class: I